

EU Declaration of Conformity

Legal Manufacturer

- Name	ABENA A/S	www.abena.com
- Address	Egelund 35	Phone: +45 7431 1818
Phone / fax / email / webpage	DK-6200 Aabenraa	Fax: +45 7462 9737
	Denmark	Mail: info@abena.com
- Single Registration Number (SRN), if applicable	DK-MF-000002482	

Medical Device(s)

- Basic UDI-DI	Please see appendix I
- Product/trade name(s) and/or product code(s)	Please see appendix I
(REF)/and or catalogue number	
- Other ref. allowing identification (e.g. UDI-DI)	Please see appendix I
- Intended Purpose	Please see appendix I
- Risk classification	Class 1 according to rule no. 1 and 5 in MDR annex
	VIII

Other information (if applicable)

- Common Specifications used for compliance	n/a
- Notified Body name and identification no. and	SATRA Technology Europe Limited
description of the conformity assessment	Bracetown Business Park
procedure performed	Clonee D15YN2P
	Ireland
- Additional information	PPE Regulation (EU) 2016/425
	EU Type Examination Module B and On-going
	conformity Module C2 performed by notified body
	2777, SATRA Technology Europe Limited.
	Issued EU Type Examination certificate number
	2777/17447-02/E01-01
- Standards used to assure compliance	Please see appendix II

The above mention manufacturer hereby declare that the above mentioned medical device(s) are compliant with the EU Regulation 2017/745 for Medical Devices and the EU legislations mentioned under "additional information".

This Declaration of Conformity is issued under the sole responsibility of the above mentioned manufacturer.

Name and Function: Khalid Elamri, Global Category Manager	
Signature	Place and date of issue
Eldestri	Aabenraa, DK, 07-12-2021
Abena, Egelund 35, DK-6200 Aabenraa	

Template Responsible: JOHO	Created: JOHO	Approved: ULDA
File: Medical-Exam-Nitril-Gloves-	Revision/Version: 1.2	Date: 16-06-2020
Classic 1999902086-2089,		
1999905271-1999905275, CN		
Copyright: Abena	Applicable SOP: n/a	Page 1/2



Appendix I, List of products

Product Name	Item no.	Basic UDI-DI	Intended Purpose
Classic Nitril Examination gloves	1999902086-		Protect patient and user
	199992090	570353800MeNF-	from cross-
powder free blue colour	1999905271-	00I-06006CU	contamination.
200 pcs and 100 pcs	1999905275		

Appendix II, List of applicable standards used

Product Name	Standards used
Nitril Examination gloves powder free	EN 455-1:2020
Nitril Examination gloves powder free	EN 455-2:2015
Nitril Examination gloves powder free	EN 455-3:2015
Nitril Examination gloves powder free	EN 455-4:2009
Nitril Examination gloves powder free	EN ISO 374-1:2016+A1 2018
Nitril Examination gloves powder free	EN 16523-1:2015+ A1:2018
Nitril Examination gloves powder free	EN ISO 374-5:2016
Nitril Examination gloves powder free	EN 374-4:2019
Nitril Examination gloves powder free	EN ISO 21420:2020

Template Responsible: JOHO	Created: JOHO	Approved: ULDA
File: Medical-Exam-Nitril-Gloves-	Revision/Version: 1.2	Date: 16-06-2020
Classic 1999902086-2089,		
1999905271-1999905275, CN		
Copyright: Abena	Applicable SOP: n/a	Page 2/2