

## EU Declaration of Conformity

According to ANNEX IV of the Medical Device Regulation (EU) 2017/745

Manufacturer:	Wenzhou K.L.F. Medical Plastics Co., Ltd No.29 Gangqiang Road, Airport New Area, Yongxing Street, Longwan District, 325000 Wenzhou, Zhejiang, People's Republic of China
SRN of the Manufacturer:	CN-MF-000011214
Authorised Representative:	Shanghai International Holding Corp. GmbH(Europe) Eiffestraße, 20537 Hamburg, Germany
SRN of the Authorised Rep.:	DE-AR-000000001
Product Name:	EXTENSION SET
Product Code:	50070,50150,50200,51070,51115,51130,51150,51155,51200.
Basic UDI-DI of Product:	6944262906B0901WB
Intended Purpose:	EXTENSION SET is intended to establish an extension of tubing where the standard length of the tubing in an intravenous administration set or transfusion set is insufficient.
EMDN Code:	A030201- EXTENSIONS

Classification (MDR, Annex VIII): **Ila, rule 2, 1<sup>st</sup> indent**

Conformity Assessment Procedure: **Annex XI**

We (manufacturer) herewith state that the above-mentioned product is in conformity with the following Medical Device Regulation, Common Specifications and Product Standards. We are solely responsible for the EU declaration of conformity.

The applicable Medical Device Regulation, Common Specifications and Product Standards:

Medical Device Regulation (EU) 2017/745

Common Specifications: N/A

**EN ISO 8536-9:2015**


Notified Body: **TÜV SÜD Product Service GmbH**  
**Ridlerstr. 65, 80339, München, Germany**

Identification number: **CE0123**

(EC) Certificate(s): **G20 047985 0029 Rev.00**

Expire date of the Certificate: **2027-12-07**

Signature:



Name: **Yuewen Jiang**

Position: **Person responsible for regulatory compliance**

Place, Date of Issue: **Wenzhou, 2023-02-23**